



**ASSOCIAZIONE ITALIANA ALLEVATORI
LABORATORIO STANDARD LATTE**

VIA DELL'INDUSTRIA 24 – 00057 MACCARESE (ROMA)
TEL. 06/6678830 FAX. 06/6678811 e mail ls@aia.it

**AFLATOXIN M1 – CERTIFIED REFERENCE MATERIAL
(SKIMMED LIOPHILIZED COW MILK)
Batch n.170407**

Sample	M1 (ppt)	SR ⁽¹⁾ (ppt)	add water to reconstitute ⁽²⁾
A	35,46	$\pm 5,36$	3,5g $\pm 0,05g$
B	46,41	$\pm 7,05$	3,5g $\pm 0,05g$
C	16,54	$\pm 2,86$	3,5g $\pm 0,05g$
D	53,56	$\pm 9,10$	3,5g $\pm 0,05g$

1) Standard deviation of Reproducibility (SR) is the measurement uncertainty of the reference material obtained from the statistical elaboration of Aflatoxin M1 Proficiency Test – April 2007

2) IT IS NECESSARY TO RECONSTITUTE THE SAMPLES WEIGHTING DEIONIZED WATER LIKE INDICATED IN THE TABLE.

EXPIRE DATE 16/06/08

Responsable of Laboratory
(Dott.ssa Annunziata Fontana)