



**ASSOCIAZIONE ITALIANA ALLEVATORI
LABORATORIO STANDARD LATTE**

VIA DELL'INDUSTRIA, 24 - 00057 MACCARESE
TEL.06-6678830 FAX 06-6678811 e mail Isl@aia.it

**CERTIFICATED REFERENCE MATERIAL
BOVINE MILK SAMPLES – UREA CONTENT
II SEMESTER 2004**

Sample	Sample Volume (ml)	Urea content (mg/100ml) (1)	Standard deviation (mg/100ml) (2)
n° 0	3	0	/
n° 1	3	24,74	± 0,61
A	24	17,91	± 0,69
B	24	30,85	± 0,82
C	24	43,96	± 0,92
D	24	52,01	± 1,15

- 1) Urea content of the reference material sample
- 2) The Standard deviation is calculated on the results from labs using differential pH instruments, participating at the Proficiency test of the 06/07/04.

Maccarese, 15/07/2004

Il Responsabile del Laboratorio

(Dr. Ugo Paggi)